




10/4 – 100 um filter change 10/12 – Analyzed QC samples, changed thw 100, 1 and 0.45 um filters 10/14 – Post filter change TOC DOC overlap 10/19 – Synced Sievers clock to computer clock, 6 min difference, changed acid cartridge in Sievers, changed 100 um filter 10/27 – Changed 100 and 1 um filters 11/2 – Changed 100 um filter 11/16 – Changed 100, 1 and 0.45 um filters, analyzed all QC samples 11/30 – Changed the 100 and 0.45 um filters, analyzed all QC samples 12/5 – Changed 100 and 1 um filters 12/8 – Replaced the supressor and guard and analytical columns on the Dionex 12/9 – Cleaned the main intake and replaced instrument drain lines



**Events:** 10/27 = Modem circuit tripped, cleaned pump intake, QC & new filters. 11/1 = Power or computer outage. 11/18 = QC & new filters, cleared vegetation from pump intake area. 11/26 – 28 = Computer outage? 11/30 – 12/2 = Power outage. QC & new filters.




**The Sievers:** 10/20-12/9 - has been operational and reporting. From 12/2-12/8 there was overlap in TOC and DOC data due to a glitch in the instrument software. The Sievers was restarted and is currently functioning normally. The overlapping data was not omitted from the graphical update because it serves as an approximate value of the true data values. **The Dionex:** 10/31-11/4 the Dionex was down because it needed a new separator and guard column. A new suppressor was installed 11/18, which significantly improved resolution of anions. The Dionex is currently functioning normally.



Significant Events: October 20<sup>th</sup> 2016 to December 9<sup>th</sup> 2016

- 10/28 RTMQC: Changed all filters
- 11/30 RTMQC: Changed all Filters



- > 10/25 YSI stopped logging data to the station computer.
- > 10/31 YSI internal data filled up and caused the recording to stop.
- > 11/22 Filter change and cleaning the YSI housing caused a shift in readings.

