


Banks Sensors



9/29: QC & filter change. 10/4: prefilter change. 10/12: QC & filter change. 10/14-17: Power outage (?) 10/19: Changed acid cartridge & prefilter




Events: 9/23 = Cleaned IOS. 9/27 = QC and filter change. 10/4 = Changed prefilters. 10/5 = Fixed flow sensor. 10/12 = QC and prefilter change.



The Sievers: 9/21-9/23 - there was a malfunction with the flow sensor causing values to overlap. 9/27-10/4 there was an acid delivery issue causing values to overlap. The Sievers is currently operational and reporting. The Dionex: 10/14-10/17 - online reporting wasn't turned back on after a QC test. New check standard was made in an effort to fix nitrate peaks. Peaks have not improved. Remediation efforts remain ongoing. The Dionex is currently operational and reporting.






Significant Events: September 23th 2016 to October 20th 2016

- 9/27 RTMQC: Changed all filters.
- **10/17 Power Outage:** The Sievers restarted itself due to a power outage, I had to manually start the instrument.

Gianelli All Sensors



- > 9/29 & 10/10 Clogging of filters and fouling of YSI housing caused questionable readings for nitrate and turbidity.
- > 10/10 to 10/14 Air got pulled into the Dionex, via the DI container, and caused the instrument to shut down.

