




5/3 to 7/17: Normal station and sample delivery system maintenance and QC over the time period. Replaced analyzer consumables on 7/17.



4/27: Power outage 5/3: New Filters, restarted the carbon analyzer and associated software to get the data to be recognized as two separate streams again. 5/4: Checked carbon analyzer and associated software to regain the TOC label for the correct data. 5/8: All data labelled DOC lately. 5/11: Switched some plugs around to see if power issue would be solved. 5/16: All new filters. Another power issue. Breaker may need replacement. 5/31: All new filters 6/12: The submersible pump motor went out yesterday. I was able to get it to start again. Changed all filters, cleaned all filter housings.




5/4 to 7/12: Normal station maintenance and QC analysis. 6/26: Anion analyzer annual PM performed. 7/20: Installed new plumbing for EXO2 sonde.



Significant Events: Apr 22^{th,} 2018 to July 20^{th,} 2018

• 5/20 Dionex Calibration: The Eluent Generator for the Dionex was swapped for a new one. The calibration performed afterwards failed. The pump wasn't diluting the calibration sample properly. The lines were flushed.



> 7/5 to 7/10 – The Dionex ran low on DI water causing it to pull air into the system. No further issues after restarting the instrument.

